Economy Pressure Indicators

Ideal Lab / Site Tool for checking & Calibration of Pneumatic Pressure / Differential Pressure Gauges / Switches / Transmitters / Receiver Gauges etc.

SALIENT FEATURES

| Digital Pressure Indicator |
|---|
| Silicon Keys |
| Selectable Engineering Pressure Units |
| DC Current Measurement |
| Switch Test capability & Zero Offset facility |
| Pressure Transducer – all Media compatible with SS316 |
| Low Battery Indication |
| Auto Power Off |
| Rugged Protection Holster |
| Ingress Protection Rating : IP 52 |
| |



MPCE-I-P

SPECIFICATIONS

| Ranges | -1 to 2 / 10 / 20 / 35 / 40 / 60 bar |
|------------------------|--------------------------------------|
| Accuracy | ±0.1% F.S. |
| Display Resolution | Minimum 4½ digits |
| Display | 7 Segment FSTN (Large) Display |
| Engineering Units (12) | Refer Table below |
| mA Input | 0 to 25 mA DC (Active) |
| Process Connection | ½" BSP (M) |
| Over Pressure Alarm | 105% FS (Audio / Visual) |
| Battery Type | 9 Volts Alkaline |

7 SEGMENT CLASSIFICATION OF ENGINEERING UNITS

| Range | Pressure / Engineering Units – Resolution | | | | | | | | | | | |
|---------|---|-------------------|--------------------|--------------------|-------|------|------|-----|-------|-------|--------------------|------|
| (bar) | inH ₂ O | mH ₂ O | cmH ₂ O | mmH ₂ O | inHg | mmHg | Mbar | kPa | Мра | bar | Kg/cm ² | psi |
| 2 | 0.1 | 0.01 | 1 | 1 | 0.001 | 0.1 | 1 | 0.1 | 0.001 | 0.001 | 0.001 | 0.01 |
| 10 | 1 | 0.1 | 1 | | 0.1 | 1 | 1 | 1 | 0.001 | 0.001 | 0.001 | 0.01 |
| 20 ~ 35 | 1 | 0.1 | | | 0.1 | | 1 | 1 | 0.001 | 0.01 | 0.01 | 0.1 |
| 40 | 1 | 0.1 | | | 1 | | | 1 | 0.001 | 0.01 | 0.01 | 0.1 |
| 60 | 1 | 0.1 | | | 0.1 | | | 1 | 0.001 | 0.01 | 0.01 | 0.1 |

STANDARD INCLUSIONS

- Pressure Indicator
- Test Leads & Seals Kit
- 9V Alkaline Battery

OPTIONAL (ADD-ON) ITEMS

- Hand Pump + Hose with End Fittings
- **Quick Coupling**
- **USB** Interface
- Rechargeable Battery with External Charger
- 'Calpress' Calibration Software with Cable
- Calibration Certificates are issued in Accordance with our Scope as granted by NABL per ISO/IEC 17025:2017 Standards
- **Carrying Case**

- Traceable Calibration Certificate
- Instruction Manual

Set of Adaptors 1/8" BSP(F) X 1/4" BSP (M) ½" BSP(F) X ¼" BSP (M) ¾" BSP(F) X ¼" BSP (M) 1/8" NPT(F) X 1/4" BSP (M) 1/4" NPT(F) X 1/4" BSP (M) 1/2" NPT(F) X 1/4" BSP (M) 3/8" NPT(F) X 1/4" BSP (M)

Manufactured by

Export Sales

NAGMAN INSTRUMENTS & ELECTRONICS (P) LTD.

Chennai-Bangalore National Highway, Nazarathpet P.O., Chembarambakkam, Chennai - 600 123. INDIA.

AN ISO 9001:2015 **CERTIFIED COMPANY**

Since 1972

Domestic Sales Phone: 044 – 66777020, 005, 021, 024. **Fax**: 044 – 66777050. **Phone**: +91-44-66777006, 008.

Fax: +91-44-66777050.

E-Mail: mktgchennai@nagman.com E-Mail: exports@nagman.com

www.nagman.com